



# Analysenzertifikat Certificate of Analysis

virion\serion

## SERION ELISA *control* ASPERGILLUS FUMIGATUS IgM

Best.-Nr. / Order No.	BR132M
Chargen-Nr. / Lot No.	SBV.CK
Verfallsdatum / Expiry Date	2015-02
OD (405 nm) Zielwert / Target Value	1,128
OD (405 nm) Zielbereich / Target Range	$\geq 0,564$ und / and $\leq 2,256$
Zielwert (Titer) / Target Value (Titre)	266,1 U/ml
Zielbereich (Titer) / Target Range (Titre)	$\geq 133,1$ U/ml und / and $\leq 532,2$ U/ml
Rekonstitution / Reconstitution	0,1 ml Aqua dest.
Serumverdünnung / Serum Dilution  Bitte RF-Absorbens verwenden! Please use Rf-absorbent!	1 + 100  Bitte Gebrauchsanweisung zu RF- Absorbens Z 4 / Z 200 beachten! Please refer to the instruction manual of Rf-absorbent Z 4 / Z 200!
Versions-Nr. / Version No.	2010 / 01-1

Hersteller / Manufacturer:  
Institut Virion\Serion GmbH  
Friedrich-Bergius-Ring 19  
D-97076 Würzburg  
Germany  
Tel. +49 931 3045 0  
Fax +49 931 3045 100  
E-Mail: [dialog@virion-serion.de](mailto:dialog@virion-serion.de)  
Internet: [www.virion-serion.de](http://www.virion-serion.de)