


# Analysenzertifikat Certificate of Analysis



virion\sersion

## SERION ELISA *control* BORDETELLA PERTUSSIS IgA

Best.-Nr. / Order No.	BR120A
Chargen-Nr. / Lot No.	SKZ.DA
Verfallsdatum / Expiry Date	2019-10
OD (405 nm) Zielwert / Target Value	1,00
OD (405 nm) Zielbereich / Target Range	$\geq 0,50$ und / and $\leq 2,00$
Zielwert (Titer) / Target Value (Titre)	57,8 IU/ml
Zielbereich (Titer) / Target Range (Titre)	$\geq 28,9$ IU/ml und / and $\leq 115,6$ IU/ml
Rekonstitution / Reconstitution	0,1 ml Aqua dest.
Serumverdünnung / Serum Dilution	1 + 100
Versions-Nr. / Version No.	2010 / 02-1

Hersteller / Manufacturer:  
Institut Virion\Serion GmbH  
Friedrich-Bergius-Ring 19  
D-97076 Würzburg  
Germany  
Tel. +49 931 3045 0  
Fax +49 931 3045 100  
E-Mail: [dialog@virion-serion.de](mailto:dialog@virion-serion.de)  
Internet: [www.virion-serion.de](http://www.virion-serion.de)