



Analysenzertifikat Certificate of Analysis

virion\serion

SERION ELISA *control* BORDETELLA PERTUSSIS IgA

Best.-Nr. / Order No.	BR120A
Chargen-Nr. / Lot No.	SMV.AH
Verfallsdatum / Expiry Date	2015-12
OD (405 nm) Zielwert / Target Value	1,463
OD (405 nm) Zielbereich / Target Range	$\geq 0,732$ und / and $\leq 2,926$
Zielwert (Titer) / Target Value (Titre)	111,0 IU/ml
Zielbereich (Titer) / Target Range (Titre)	$\geq 55,5$ IU/ml und / and $\leq 222,0$ IU/ml
Rekonstitution / Reconstitution	0,15 ml Aqua dest.
Serumverdünnung / Serum Dilution	1 + 100
Versions-Nr. / Version No.	2010 / 02-1

Hersteller / Manufacturer:
Institut Virion\Serion GmbH
Friedrich-Bergius-Ring 19
D-97076 Würzburg
Germany
Tel. +49 931 3045 0
Fax +49 931 3045 100
E-Mail: dialog@virion-serion.de
Internet: www.virion-serion.de