



Analysenzertifikat Certificate of Analysis

virion\serion

SERION ELISA *control* ENTEROVIRUS IgG

Best.-Nr. / Order No.	BR133G
Chargen-Nr. / Lot No.	SAX.FD
Verfallsdatum / Expiry Date	2015-12
OD (405 nm) Zielwert / Target Value	0,951
OD (405 nm) Zielbereich / Target Range	$\geq 0,476$ und / and $\leq 1,902$
Zielwert (Titer) / Target Value (Titre)	98,0 U/ml
Zielbereich (Titer) / Target Range (Titre)	$\geq 49,0$ U/ml und / and $\leq 196,0$ U/ml
Rekonstitution / Reconstitution	0,10 ml Aqua dest
Serumverdünnung / Serum Dilution	1 + 250
Versions-Nr. / Version No.	2010 / 01-1

Hersteller / Manufacturer:
Institut Virion\Serion GmbH
Friedrich-Bergius-Ring 19
D-97076 Würzburg
Germany
Tel. +49 931 3045 0
Fax +49 931 3045 100
E-Mail: dialog@virion-serion.de
Internet: www.virion-serion.de