

Institut Virion\Serion GmbH Würzburg

ZERTIFIKAT QUALITÄTSKONTROLLE


CERTIFICATE OF QUALITY CONTROL
CERTIFICADO DI CONTROLLO DI QUALITA

SERION ELISA classic:	COXIELLA BURNETII PHASE 1 IgG KIT
Ch.-B./Lot No.:	SHZ.BS
Verfallsdatum/Exp.date:	2011-02

Kontrollseren Control sera	Ch.-B. Lot No.	gemessener Wert measured value OD 405 nm	Gültigkeitsbereich validity range
Positivkontrolle Positiv control	SHZ.BQ	1,09	> 0,55
Cut off-Serum cut off serum	SHZ.BR	0,52	0,26 – 0,78

Teststreifen/strips:	COXIELLA BURNETII PHASE 1 IgG
Ch.-B./Lot No.:	SBZ.AN
Konjugat/conjugate	IgG +++
Ch.-B./Lot No.:	SKY.BH

27.08.2009
Prüfdatum


Möhlig, Heike
Testbetreuer / Product Specialist