

Institut Virion\Serion GmbH Würzburg

ZERTIFIKAT QUALITÄTSKONTROLLE

CERTIFICATE OF QUALITY CONTROL
CERTIFICADO DI CONTROLLO DI QUALITA

| | |
|--------------------------------|--------------------------------------|
| SERION ELISA classic: | COXIELLA BURNETII PHASE 2 IgM Kit |
| Ch.-B./Lot No.: | SFZ.BB |
| Verfallsdatum/Exp.date: | 2011-03 |

| Kontrollseren Control sera | Ch.-B. Lot No. | gemessener Wert measured value OD 405 nm | Gültigkeitsbereich validity range |
|---|---------------------------|---|--|
| Positivkontrolle Positiv control | SFZ.AZ | 0,80 | > 0,40 |
| Cut off-Serum cut off serum | SFZ.BA | 0,47 | 0,24 – 0,71 |

| | |
|-----------------------------|-------------------------------|
| Teststreifen/strips: | Coxiella burnetii Phase 2 IgM |
| Ch.-B./Lot No.: | SEZ.DC |
| Konjugat/conjugate | IgM + |
| Ch.-B./Lot No.: | SLY.BT |

15.06.2009

Prüfdatum



Dr. Andreas Schmiiedl
Testbetreuer / Product Specialist